How Do I Set the Software to be Compliant for Canada?

Last Modified on 09/27/2018 8:57 am PDT

This article will show you how to change your software to be Canada compliant after it has been installed.

1. Open the Setup menu, go down to Settings, and select General from the list on the left.



2. Click the Advanced button. Click to remove the check mark in the "Measure Mode" box to make it compliant for Canadian laws. Click OK to save your settings.

General Setup	x						
Show training prompt on startup K							
Show add/replace exam prompt Cance	el						
Check for and use internet services if available							
Confirm successful hub connection on startup							
Load libraries on program startup							
Show enhanced Patient Exam information	ced						
Mute MyoVision Software sounds							
Display Battery Voltage in Device Status							
General Setup	x						
Show training prompt on startup							
Remember last patient OK							
Show add/replace exam prompt Cance	el						
Check for and use internet services if available							
Confirm successful hub connection on startup							
Load libraries on program startup							
Show enhanced Patient Exam information	8						
Mute MyoVision Software sounds							
Display Battery Voltage in Device Status							
_ Video Path Previous Fxam History:							
Path C:\WinScan98\Training2							
Exam Goal Display Timer — Enhanced GUI 📃 Use SQL re	mote data						
Measure Mode (USA)	rt Disclaimer						
Force Lheck Device Automatic D	ану васкир						
Seconds: Status I Bratient ID Required I Hide Legacy	y Protocols						
FlexUltra Only Mode Disable Visio	onPort 2						
Debug Save Perfor	med Exams						
RESET Primary user Info. Enable Low Battery Pop-ups							

3. Open the Setup menu, go down to Settings, and select Static.

🍘 MyoVision (tm)									
File	View	Patient	Exam	Protocol	Set	up Help		_	
	L	عن	\bigotimes	<u>ل</u> ع		Doctor Information			\Leftrightarrow
Prin		Open	Close	Backup		Database		uick	Training
		open				Calibrate	►		
					-	Settings	•		General
						Reports	≁		Devices
						Language			Pair Device \ Check Status
									VISIONPOR
									Static
									Dynamic

4. Click the "EP Score" button. Under the area marked "Score Settings," uncheck the box next to "Enable." click OK to save your Settings.

Static Settings X	Static Settings X
Static Settings × Bar Options: Very Low Very Low Goal Imbalance: Show Goal Imbalance: Show Site Labels Imbalance: Show Site Labels Imbalance: Show p.V Moderate MWA High Show p.V High Moving Bar Maximizer Important Countdown Darker Lighter EP Score >>	Bar Options: Very Low Goal Percent Very Low Goal Show Goal Preview Low Imbalance: 3D Bars Default Colors Low Imbalance: Show Site Labels Default Colors Moderate MWA Hide Bar Pointer Tip Cancel Moderate MWA Hide Bar Pointer Tip Cancel Moderate MWA Backgraph Image Cancel Wight Darker Lighter < <less< td=""> Display exam countdown Darker Lighter <<less< td=""> Score Settings Display Size: Medium Show Score value when overlaying exams only. Text: EP Stress Score Color Single Mode Compare Mode Current: 255=0paque 200</less<></less<>

The software is now set to be compliant for use in Canada.